


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-MV-1-C-P*

S.N.: *5868*

Date: *MAY 22 1991*



*60 db atten*

*10.0 VOLTS/DIV*

*PRF = 5 KHz.*

- a) Output signal amplitude:  
*0 TO +100 VOLTS*
- b) Pulse width:  
*3 TO 100 NS*
- c) Rise time:  
*≤ 1 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz.*

*[Signature]*