


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-651-EA-7D

S.N.: 5872

Date: JUNE 26 1991



60 db ATTEN

100 VOLTS DIV

PRF = 10 KHz

a) Output signal amplitude:

0 TO 100 VOLTS

b) Pulse width:

10 NS

c) Rise time:

$\leq$  1 NS

d) Fall time:

$\leq$  2 NS

e) PRF:

0 TO 20 KHz

f) Jitter, stability:

OK

g) Prime power:

+15 V

250 mA (max)

*[Signature]*