


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVN-3-C-D*

S.N.: *5874*

Date: *JUNE 14 1991*




- a) Output signal amplitude:  
*0 TO +150V (f < 100 MHz)*
- b) Pulse width:  
*0 TO +5 V (f = 250 MHz)*

- c) Rise time:  
*0.2 TO 1.0 NS*
- d) Fall time:  
*≤ 150 PS*

- e) PRF:  
*50 TO 250 MHz*
- f) Jitter, stability:

- g) Prime power:  
*OK*  
*120/240V*  
*50-60 Hz*

Ⓐ *4066 ATTEN*  
*10 DIV/10V*  
*PRF = 100 MHz*



Ⓑ *4066 ATTEN*  
*5 DIV/10V*  
*PRF = 200 MHz*

*[Signature]*