


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U-4V2-P-CEA*

S.N.: *5878*

Date: *MAY 2 1991*



60 dB ATTEN

100 VOLTS/DIV

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
≤ 8 TO 20 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS
250 mA (MAX)

[Signature]