


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-1-PW-LP*

S.N.: *5085*

Date: *MAY 17 1991*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
0.1 TO 5.0 μ S
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V

200 NS/DIV

50 VOLTS/DIV

$R_L = 50 \Omega$

PRF = 1.0 KHz

50-60 Hz