


PULSE GENERATOR

PERFORMANCE CHECK


Model: *PAH-141-PN-SSB*

S.N.: *5897*

Date: *JUNE 6 1991*



*Point 60db ATTEN  
100 VOLTS/DIV*



*Point 60db ATTEN  
100 VOLTS/DIV*

- a) Output signal amplitude:  
*P : 700 VOLTS  
N : -100 VOLTS*
- b) Pulse width:  
*0.6 TO 2.0 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VOLTS  
250 mA (MAX)*