


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APR-1-PW-CP-MA2*

S.N.: *5900*

Date: *MAY 31 1991*



- a) Output signal amplitude:
0 TO +150 VOLTS
- b) Pulse width: *TO 6.5*
0.3 TO 2.0 μ S
- c) Rise time: *(1% MAX DUTY CYCLE)*
 ≤ 50 NS
- d) Fall time:
 ≤ 50 NS
- e) PRF: *20 KHZ TO 20 KHZ*
(1% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:

R_c = 6 Ω
50 VOLTS/DIV
300 NS/DIV
PRF = 5 KHZ
(MAX DUTY CYCLE)

120/240 V
50-60 HZ