


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-1-C-P-OT*

S.N.: *5902*

Date: *JUNE 5 1991*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTEN

5. VOLTS DIV

PRF ≈ 12 MHz

120/240V
50-60 Hz
[Signature]