


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVH-MU1-PW-55C*

S.N.: *5915*

Date: *JULY 10 '91*



*Ⓐ* *POUT 60 db ATTEN*



*Ⓑ* *POUT 60 db ATTEN*  
*PRF = 5 kHz*

- a) Output signal amplitude:  
*± 100 VOLTS*
- b) Pulse width:  
*0.6 TO 2.0 ns*
- c) Rise time:  
*ns*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 100 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+10.2 TO 14.0 VOLTS*  
*250 mA (MAX)*

*[Signature]*