


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVH-1741-PN-55C

S.N.: 5916

Date: JUL 7 10 1991



(A) P_{out} 60 dB ATTEN



(B) N_{out} 60 dB ATTEN
50 kHz 100F

a) Output signal amplitude:

± 100 uV

b) Pulse width:

0.6 TO 2.0 ns

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 100 kHz

f) Jitter, stability:

OK

g) Prime power:

+ 10.2 TO 14.0 V

250 mA (MAX)

[Signature]