



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-2-C-PN*  
S.N.: *5922*  
Date: *JULY 18 1991*



Ⓐ *40 dB ATTEN*  
*10 VOLTS/DIV*  
*PRF = 700 KHz*



Ⓑ *AS Ⓐ BUT*  
*1 NS/DIV*  
*(RISE TIME)*

- a) Output signal amplitude:  
*0 TO 770 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 VOLTS*  
*50-60 MHz*

*[Signature]*