


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV6-1-P

S.N.: 5926

Date: JUNE 4 1991



60 dB ATTEN

100 VOLTS/DIV

PRF = 20 KHz

- a) Output signal amplitude: 0 TO +250 V
- b) Pulse width:  $\leq 4$  NS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 50 KHz
- f) Jitter, stability: OK
- g) Prime power: +15 VOLTS, 250 mA (max)

4/6