


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1702-C-P*

S.N.: *5928*

Date: *JUNE 13 1991*



40 dB ATTEN

10 VOLTS/DIV

PRF = 100 KHz

- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
0.3 TO 2.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PS
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

1.20 / 240 V
50-60 KHz