


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-14V.2-P-00A*

S.N.: *5929*

Date: *JUNE 28 1991*



60 db ATTEN

100 VOLT/DIV

PRF = 20 kHz

- a) Output signal amplitude:
0 TO 100 VOLT
- b) Pulse width:
0 TO 2 NS
- c) Rise time:
± 2 NS
- d) Fall time:
± 2 NS
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 V DC

250 mA (CH 1A)