


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AWH-141-C-SP1A-M-P*

S.N.: *5931*

Date: *JUNE 18 1991*



- a) Output signal amplitude: *0 TO 1100 V.D.M.S*
- b) Pulse width: *0.6 TO 2.0 NS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz

60 db ATTEN
100 VOLTS/DIV
100 KHz

[Signature]