


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2A-C-PN

S.N.: 5940

Date: JULY 17 1991



60 dB ATTEN

≈ 100 VOLTS/DIV

PRF = 5.0 KHz

a) Output signal amplitude:
0 TO ± 176 VOLTS

b) Pulse width:
5 TO 100 NS

c) Rise time:
≤ 2 NS

d) Fall time:
≤ 3 NS

e) PRF:
0 TO 5 KHz

f) Jitter, stability:
OK

g) Prime power:
120/240 V
50-60 Hz