


PULSE GENERATOR

PERFORMANCE CHECK


Model: *PUL-AU-1-C-W-PN-05-M*

S.N.: *5941*

Date: *JUNE 25 1991*



*PENT 60 db ATTEN
100 VOLTS/DIV
(RISE TIME)*



*PENT 60 db ATTEN
100 VOLTS/DIV*

- a) Output signal amplitude: *0 TO ± 100 VOLTS*
- b) Pulse width: *5 TO 400 NS**
- c) Rise time: *≤ 1 NS*
- d) Fall time: *≤ 3 NS*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V
50-60 Hz*