


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVO-1-C-P

S.N.: 5943

Date: JULY 22 1991



≈ 60 AMP/Div

PRF = 500 Hz

- a) Output signal amplitude:
+25 TO +100 AMP
- b) Pulse width:
≈ 25 NS
(1 HALF POWER)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz