


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-402-P-CEA*

S.N.: *5944*

Date: *SEPT 30 1991*



60 db ATTEN

100 VOLT/DIV.

PRF = 20 kHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
0 TO 20 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
250 mA (MAX)