


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-M-1402-PS-PN*

S.N.: *5953*

Date: *July 23 1991*



40 dB ATTEN
10 VOLTS/DIV
PRF = 100 KHz

- a) Output signal amplitude:
0 TO ± 20 VOLTS
- b) Pulse width:
0.3 TO 2.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 200 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 MHz

[Signature]