


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APX-FD2-PS-IP*

S.N.: *5954*

Date: *OCT 11 1991*



a) Output signal amplitude: *TTL TO 50m*

b) Pulse width: *10 NS TO 10 US*

*IN*  
c) Rise time: *NA*

*OUT*  
d) Fall time: *NA*

e) PRF: *100 Hz TO 100 MHz*

f) Jitter, stability: *OK*

g) Prime power: *120/240V 50-60 Hz*

h) DIVISION FACTOR: *2 TO 255.*

*DIVISION FACTOR = 17.  
PRF<sub>in</sub> ≈ 50 MHz.*