


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-3A-D-LLA*

S.N.: *5959*

Date: *JULY 16 1991*



*70 db ATTEN*

*≈ 320 VOLTS/DIV*

*PRF ≈ 10 kHz*

a) Output signal amplitude:

*≈ 300 VOLTS*

b) Pulse width:

*10 NS*

c) Rise time:

*≤ 600PS*

d) Fall time:

*≤ 2 NS*

e) PRF:

*0 TO 20 kHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 VOLTS*

*250 mA (MAX)*

*[Signature]*