


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ACL-3A-P-LLA*

S.N.: *5960*

Date: *JUL 16 1991*



*70 db ATTEN
= 320 VOLTS/DIV
PRF = 10 KHz*

- a) Output signal amplitude:
≥ 300 VOLTS
- b) Pulse width:
10 NS
- c) Rise time:
≤ 600 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*+ 15 VOLTS
250 mA (max)*

[Signature]