


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MO-P.*

S.N.: *5976*

Date: *JULY 29 1991*



- a) Output signal amplitude:
+10 TO +50 AMP)
- b) Pulse width:
≈ 20 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
250 mA (MAX)

≈ 20 AMP/ DIV
PRF = 1.0 KHz