


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-1-T*

S.N.: *5977*

Date: *JULY 30 1991*



60 db ATTEN

100 VOLTS/DIV

PRF = 40 KHz

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
≤ 4 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLTS

250 mA (MAX)

[Signature]