


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-1-P*

S.N.: *5978*

Date: *JULY 30 1991*



*60 db ATTEN*

*100 0.2 V/div*

*PRF ≈ 40 KHz*

- a) Output signal amplitude: *0 to +250 VOLTS*
- b) Pulse width: *≤ 4 NS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 50 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VOLTS*  
*250 mA (MAX)*