


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-MP-P*

S.N.: *5979*

Date: *SEPT 19 1991*



50dB ATTEN

32 VOLTS/DIV

FBF = 50 kHz

- a) Output signal amplitude:
71 ± 40 VOLTS
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
≤ 350 ps
- d) Fall time:
≤ 2 NS
- e) PRF:
0-1 MHz (PW < 2 NS)
0-20 kHz (PW = 100 NS)
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS DC
150 mA (MAX)