


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-3A-P-1M1*

S.N.: *5985*

Date: *JULY 31 1991*



- a) Output signal amplitude:
0 TO +35 AMP
- b) Pulse width:
FWHM < 5 NS
FW0.1M ≤ 10 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 10 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

~ 20 AMP 5 PDW
PKF = 3 KHZ

+ 12 VOLTS
250 mA