


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMH-3.-C-M-PN*

S.N.: *5997*

Date: *AUG 29 1991*



50 dB ATTEN

32 VOLTS/PIV

PRF ≈ 12 MHz.

- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
≤ 600 PS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50-60 Hz.

[Signature]