


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVK-AV-C-N*

S.N.: *5999*

Date: *Jul 21 1991*



*ⓐ 40db ATTEN
10 VOLTS/DIV
PRF ≈ 20 KHz*



*ⓑ AS ⓐ BUT
500 PS/DIV
(RISE TIME)*

- a) Output signal amplitude:
0 TO -15V
- b) Pulse width:
2 TO 50 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

[Signature]