


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-C-P*
S.N.: *6012*
Date: *SEPT 6 1991*



- a) Output signal amplitude:
0 TO +350 V
- b) Pulse width:
(TO 50ns)
5 TO 100 ns
- c) Rise time:
≤ 2 ns
- d) Fall time:
≤ 3 ns
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

70 db ATTEN
3320 VOLT DIV
PRF = 2 kHz