


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVI-V-14.3-C-P-OT-CA-EN

S.N.: 6021

Date: SEPT 12 1991



Ⓐ 60 db ATTEN
100 VOLTS/DIV
PRF = 5 KHz



Ⓑ AS Ⓐ BUT
1 NS/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO 1700 VOLTS
- b) Pulse width:
4 TO 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz
- h) OT: 0 TO ± 5 V
TO 50 Ω