


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-B3-CA-P*
S.N.: *6025*
Date: *SEPT 10 1991*



50 db ATTEN
≈ 32 VOLTS/DIV
PRF = 20 KHz
R_L = 50Ω

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: *(to 50ns)*
20 TO 200 NS
- c) Rise time:
≤ 3 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 VOLTS
300 mA (max)

[Signature]