


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MHS-1-C-P-TE01*

S.N.: *6032*

Date: *DEC 9 1991*



40 db ATTEN

10 VOLTS / DIV

PRF = 2.5 MHz

- a) Output signal amplitude: *0 TO 7.13 VOLTS*
- b) Pulse width: *TO 50N*
- c) Rise time: *≤ 130 PS*
- d) Fall time: *NA*
- e) PRF: *4 KHz TO 2.5 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz