


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AD-7E-C-A-ET-6W-901*


S.N.: *6039*

Date: *OCT 22 1991*



- a) Output signal amplitude:
0 TO +15 AMPS TO
- b) Pulse width: *(0 TO +20 VOLTS)*
2 µS TO 20 mS
- c) Rise time: *(10% MAX DUTY CYCLE)*
≤ 100 NS
- d) Fall time:
≤ 100 NS
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability: *(10% MAX DUTY CYCLE)*
OK

② $R_L = 1.0 \Omega$
 $\therefore 10 \text{ AMPS/DIV}$
 $\text{PRF} \approx 1 \text{ KHz}$



- g) Prime power:
120/240 V
50-60 Hz

③ $R_L = 1.0 \Omega$
 $\therefore 10 \text{ AMPS/DIV}$
 $\text{PRF} \approx 10 \text{ Hz}$