


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVAMP-3-P

S.N.: 6040

Date: SEPT 20 1991



(A) 40 dB ATTEN  
10 VOLTS/DIV  
PRF = 100 KHz



(B) 15 dB ATTEN  
1 NS/DIV  
(RISE TIME)

- a) Output signal amplitude: 0 TO +20 VOLTS
  - b) Pulse width: 8 TO 100 NS
  - c) Rise time:  $\leq 200$  PS
  - d) Fall time:  $\leq 300$  PS
  - e) PRF: 0 TO 1 MHz
  - f) Jitter, stability: OK
  - g) Prime power: +24 VOLT  
300 mA (max)
- [Signature]*