


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-AV-1-C-N

S.N.: 6052

Date: OCT 4 1991



60 db ATTEN

100 VOLT/DIV

a) Output signal amplitude:

0 TO -100 VOLTS

b) Pulse width:

3 TO 100 NS

c) Rise time:

$\leq 1$  NS

d) Fall time:

$\leq 2$  NS

e) PRF:

0 TO 5 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240V

50-60 Hz