


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-142-P-CEA*

S.N.: *6054*

Date: *NOV 27 1991*



60 db ATTEN

100 VOLTS/DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
≤ 5 ns TO 20 ns
- c) Rise time:
≤ 2 ns
- d) Fall time:
≤ 2 ns
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS.
250 mA (MAX)