


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-5A-P-M3*

S.N.: *6077*

Date: *NOV 5 1991*



- a) Output signal amplitude:
0 to +28 Amp
- b) Pulse width: *to 3 Ω*
20 to 100 ns
- c) Rise time:
≤ 8 ns
- d) Fall time:
≤ 8 ns
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:

R_L = 3 Ω

PRF ≈ 5 kHz

*+24 VOLTS,
250 mA (MAX)*

SB