


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AV100N-1-C-M-P-PN*

S.N.: *6079*

Date: *NOV 12 1991*



① *40db ATTEN*
5 VOLTS/DIV
PRF ≈ 10 MHz



② *AS* ③ *DUT*
PRF ≈ 50 MHz

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.3 TO 5.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 200 PS
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz