


PULSE GENERATOR

PERFORMANCE CHECK


Model: *PR-4B-PS-P-EAD-PAS/*

S.N.: *6091*

Date: *DEC 12 1991*



Ⓐ *HIGH DUTY CYCLE*
 $f_{re} = 100K$



Ⓑ *LOW DUTY CYCLE*
(RISE TIME)

- a) Output signal amplitude:
0 TO 300 VPP
- b) Pulse width: *TO $R_L > 100K$*
- c) Rise time: *1 μ S TO 1 μ S*
(50% WITH PULSE CLIP)
- d) Fall time: *≤ 30 NS*
- e) PRF: *≤ 30 NS*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

NY
8