


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM14-3-P*

S.N.: *6092*

Date: *NOV 20 1991*



*51 db return*

*≈ 32.00 V / DIV*

*PRF ≈ 10 MHz*

- a) Output signal amplitude:  
*0 TO +50 VOLTS*
- b) Pulse width: *TO 50N*  
*≤ 600 ps*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 VOLTS.*  
*500 mA*  
*(MAX)*

*[Signature]*