


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AK-B3-C-W-ET-EW-AT-P*

S.N.: *6094*

Date: *NOV 13 1991*



(A) $R_L = 50 \Omega$
PRF = 5.0 KHz



(B) $R_L = 50 \Omega$
PRF = 5.0 KHz

a) Output signal amplitude:

0 TO +100 V_{out}

b) Pulse width: *TO 50 ns*

0.1 TO 5.0 μ S

c) Rise time:

≤ 5 ns

d) Fall time:

≤ 5 ns

e) PRF:

0 TO 5 KHz.

f) Jitter, stability:

OK

g) Prime power:

120/240 V
50-60 Hz.
ff