


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMM4-3-C-P*
S.N.: *6101*
Date: *NOV 25 1991*



- a) Output signal amplitude: *0 TO +50 VOLTS*
- b) Pulse width: *(TO 50ns)*
0.6 ns
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz

50 dB ATTEN

~ 32 VOLTS/DIV

PRF ~ 12 MHz

120/240 V
50-60 Hz
[Signature]