


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2-C-PN-M-05

S.N.: 6116

Date: JAN 31 1992.



70 db ATTEN
≈ 320 VOLTS/DIV
PRF = 10 KHz.

- a) Output signal amplitude:
0 TO ±350 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Handwritten signature]