


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVD-9B-C-P-CAEW-MI


S.N.: 6135

Date: JAN 21 1992



- a) Output signal amplitude:
0 TO + 200 mA
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 200 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 280 V
50 - 60 Hz

Ⓐ AVD-9B-C MAIN FRAME
TO 40db ATTEN
10 VOLTS/DIV
PRF = 1.0 MHz



Ⓑ AVX-51 MI OUTPUT
TO 20db. IN 459A
BIOLOG BETWEEN PINS
1+2. 200 mA/DIV.

[Handwritten signature]