PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-1-P

S.N.: 6139

Date: JAN 22 1992



60 db ATTEN

100 VOLTS/DIV

PRF = 30 KHZ

a) Output signal amplitude:
0 TO +250 VOLTS

b) Pulse width:
≤ 4 NS

c) Rise time:
NA

d) Fall time:
NA

e) PRF: 0 TO 50 KHZ

f) Jitter, stability:
OK

g) Prime power:
+15 VOLTS
250 mA MAX