


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-1-P*
S.N.: *6141*
Date: *JAN 22 1992*



60 db ATTEN
100 VOLTS/DIV
PRF = 30 KHz

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
≤ 4 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
250 mA (max)

[Signature]