


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-1-C-P*
S.N.: *6145*
Date: *JAN 2 1992*



60 dB ATTEN
10.0 VOLTS/DIV
PRF = 20 kHz

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
≤ 4 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]