


PULSE GENERATOR
PERFORMANCE CHECK

Model: AN-143 - IBM A

S.N.: 6146

Date: JAN 19 91



- a) Output signal amplitude:
0 TO +25 VOLTS
- b) Pulse width: TO $R_L \geq 1K$
100 NS TO DC
- c) Rise time:
 ≤ 50 NS
- d) Fall time:
 ≤ 50 NS
- e) ~~PRES~~
GAIN: X5 VOLTAGE
- f) Jitter, stability:
OK
- g) Prime power:
+24 V,
350 mA MAX
- h) R_{out} : 50 Ω .

V_{out}
 $R_L = 1K$
 $V_{in} = +5$ VOLTS.